
Document Title: **EU Declaration of Conformity of 2017/745 (MDR)**

Document Number: LP-CE-LFR-01-003

Revision: V1.1

Author: Taoying Li

Date: 2020-05-18

| | Department | Superintendent | Date |
|-------------|-------------------|---------------------|------------|
| Written by | | <u>Tao Ying Li</u> | 2020-05-18 |
| Reviewed by | R&D | <u>Huang Peng</u> | 2020-05-18 |
| | Quality Assurance | <u>Xiaolong Shu</u> | 2020-05-18 |
| | Operation | <u>Bo. Ge</u> | 2020-05-18 |
| Approved | | <u>Qingjun Yi</u> | 2020-05-18 |

REVISION STATUS:

| Version | Brief Description of Revision | Author | Date (<i>DD MMM YYYY</i>) |
|---------|---|------------|-----------------------------|
| V1.0 | First procedure | Taoying Li | 2020-04-08 |
| V1.1 | Add the intended use and basic UDI-DI according to audit opinion. | Taoying Li | 2020-05-18 |

EU Declaration of Conformity

Manufacturer: Shenzhen LEPU Intelligent Medical Equipment Co., Ltd.

Address: North side of floor 3, BLD 9 BaiWangxin High-Tech Industrial Park Songbai Road, Xili Street, Nanshan District 518055 Shenzhen, Guangdong, P.R.CHINA.

Tel.: +86 0755-86952278 **Fax:** +86 0755-86952278

Website: <http://lepucare.com/>

SRN: To be registered

European representative: Lepu Medical (Europe) Cooperatief U.A.

Address: Abe Lenstra Boulevard 36, 8448 JB, Heerenveen, The Netherlands

Tel: +31-515-573399 **Fax:** +31-515-760020

SRN: To be registered

Product: Infrared forehead thermometer

Brand name: /

Intended use: Infrared forehead thermometer is an infrared thermometer intended for the measurement of human body temperature in people of all ages without contact to the body and may be used by medical professionals or by consumers in a home environment.

Basic UDI-DI: To be applied

Device Nomenclature Code: V03010102

Model List: See Annex 1

Applied Standards List: See Annex 2

Classification: According to Annex VIII, Rule 10 of (EU) 2017/745 (MDR), the Infrared forehead thermometer is in class IIa.

Conformity Assessment Route: Chapters I and III of Annex IX, and including an assessment of the technical documentation as specified in Section 4 of that Annex, Regulation (EU) 2017/745 (MDR)

We hereby declare that the above mentioned product meet the provisions of the Regulation (EU) 2017/745 (MDR) for medical devices. No medicinal product, including a medicinal product derived from human blood or human plasma, no tissues or cells of human origin or their derivatives, no CMR or endocrine-disrupting substances are incorporated into the device. All supporting documentation is retained under the premises of the manufacturer and Notified Body 2797, BSI Group The Netherlands B.V., Say Building, John M. Keynesplein 9, 1066 EP, Amsterdam, Netherlands.

CE 2797

| Certificate | Initially issued | Last renewal | Valid until |
|--|------------------|--------------|-------------|
| Full Quality Assurance System Certificate No.: | \ | \ | \ |

The EU Declaration of Conformity is issued under the sole responsibility of the manufacturer:









Shenzhen LEPU Intelligent Medical Equipment Co., Ltd.






Signed for and on behalf of :

Name : Qingjun Yi
Function (Company) : Management Representative
Date : _____ DD/MM/YYYY
Location : Shenzhen

Annex 1 Catalogue Number List

Table 1 Specifications of Infrared forehead thermometer

| Model | Structure | Button | Size (length*width*height) | Power supply | Low voltage alarm | Mode | Automatic power off/standby | Memory group | Appearance |
|--------|-----------------------------|---|----------------------------|--------------|---|---|--|--------------|--|
| LFR30B | LFR30 series main structure | <ul style="list-style-type: none"> ● On/Off Button ● Memory button ● Mode slide button | 168.5mm*36mm*48mm | d.c.3V | When the voltage of LFR30B is lower than $2.5 \pm 0.1V$, the low-voltage symbol will appear after turning on, and it cannot be measured. The low-voltage symbol is “▼” | Body mode “  ” Calibration mode “  ” | The device is automatically turned off after $1\text{min} \pm 10\text{s}$ when there is no operation | 99 groups |  |
| LFR50 | LFR50 series main structure | <ul style="list-style-type: none"> ● On/Off Button ● Memory button ● Mode button | 132mm*42mm*173.5mm | d.c.9V | When the voltage of LFR50/60 is lower than $5.8 \pm 0.2V$, the low-voltage symbol will appear after turning on, and it cannot be measured. The low-voltage symbol is “  ” | Body mode “  ” Calibration mode “  ” Room mode “  ” | The device automatically enters into standby status after 60s when there is no operation, and no further operation, it will shut down automatically in 60 minutes. | 99 groups |  |

| Model | Structure | Button | Size (length*width*height) | Power supply | Low voltage alarm | Mode | Automatic power off/standby | Memory group | Appearance |
|---------|-----------------------------|--|-------------------------------|------------------------|--|---|--|--|---|
| LFR60 | LFR60 series main structure | <ul style="list-style-type: none"> ● On/Off Button, ● Memory button ● Mode button | 129mm*41.5mm*171mm | d.c.9V | When the voltage of LFR50/60 is lower than $5.8\pm 0.2V$, the low-voltage symbol will appear after turning on, and it cannot be measured. The low-voltage symbol is “  ” | Body mode “  ” Calibration mode “  ” Room mode “  ” | The device automatically enters into standby status after 60s when there is no operation, and no further operation, it will shut down automatically in 60 minutes. | 99 groups |  |
| Remarks | Different main structure | The LFR30B switches mode by sliding the mode slide button, and the LRF50/60 switches mode by pressing down the mode button | Different size | Different power supply | Due to the different power supply, the low voltage alarm is different, and the low-power symbol is also different. | LFR50/60 are equipped with an additional room mode compared with LFR30B, and the symbol of calibration mode and body mode is also different | LFR50/60 are equipped with an additional standby function compared with LFR30B. | The memory group of all three models is the same, but LFR50/60 will display the group number on the LCD in real time, while the LFR30B will not. | Different appearance |

Accessories: There are two 1.5V AAA dry batteries packaged with the LFR30B, and one 9V dry battery packaged with LFR50/60.

Annex 2 Applied Standards/Common Specifications (CS) List

The standards/Common Specifications (CS) applicable for this product are listed as below:

| Standard /Common Specifications (CS) No. | Standard/Common Specifications (CS) Name | Date of Issue | Full/Partial Compliance |
|--|---|---------------|-------------------------|
| EN 1041:2008+A1: 2013 | Information supplied by the manufacturer of medical devices | 2013-09-25 | Full |
| EN ISO 14971:2012 | Medical devices – Application of risk management to medical devices | 2012-07-31 | Full |
| EN ISO 15223-1:2016 | Medical devices - Symbols to be used with medical device labels, labelling and information to be supplied - Part 1: General requirements | 2016-11-30 | Full |
| EN ISO 13485: 2016+AC:2018 | Medical devices - Quality management systems-Requirements for regulatory purposes | 2018-03-28 | Partial Compliance |
| IEC 62366-1:2015 | Medical devices - Part 1: Application of usability engineering to medical devices | 2015-02 | Partial Compliance |
| EN 60601-1-6:2010+A1:2015 | Medical electrical equipment - Part 1-6: General requirements for basic safety and essential performance - Collateral standard: Usability | 2015-07 | Full |
| IEC/TR 62366-2:2016 | Medical devices - Part 2: Guidance on the application of usability engineering to medical devices | 2016-04-27 | Technical guidance |
| IEC 62304:2015 | Medical device software - Software life-cycle processes | 2015-06 | Partial Compliance |
| IEC 60601-1:2005+A1:2012 | Medical electrical equipment - Part 1: General requirements for basic safety and essential performance | 2012-08 | Partial Compliance |
| IEC 60601-1-2:2014 | Medical electrical equipment - Part 1-2: General requirements for basic safety and essential performance - Collateral standard: Electromagnetic compatibility - Requirements and tests | 2014-02 | Full |
| IEC 60601-1-11:2015 | Medical electrical equipment - Part 1-11: General requirements for basic safety and essential performance – Collateral Standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment | 2015-01 | Partial Compliance |
| ISO 80601-2-56:2017+AC:2018 | Medical electrical equipment —Part 2-56: Particular requirements for basic safety and essential performance of clinical thermometers for body temperature measurement | 2018-11 | Partial Compliance |
| EN ISO 14155:2011+AC:2011 | Clinical investigation of medical devices for human subjects - Good clinical practice | 2011-07-15 | Full |
| MEDDEV 2.7/1 rev.4 | CLINICAL EVALUATION: A GUIDE FOR MANUFACTURERS AND NOTIFIED BODIES UNDER DIRECTIVES 93/42/EEC and 90/385/EEC | 2016-06 | Full |
| MEDDEV 2.12/1 rev.8 | GUIDELINES ON A MEDICAL DEVICES VIGILANCE SYSTEM | 2013-01 | Full |
| / | Additional Guidance Regarding the Vigilance System as outlined in MEDDEV 2.12-1 rev. 8 | 2019-06 | Full |
| ISTA 2A-2011(2012) | International Safe Transport Association Transport Test Standard | 2012-01 | Full |
| IEC 62506:2013 | Methods for product accelerated testing | 2013-06 | Partial Compliance |